


Centaurus A



Steinle (AIP Conf. Proc. 587, 353–357, 2001)

Broad-band spectrum of Cen A:
Spectral Energy Distribution (SED)

νf_ν is flat

⇒ similar energy output at all wavebands!

Shown is a νf_ν plot, where ν : frequency, f_ν : flux density at frequency ν (units of f_ν are $\text{J s}^{-1} \text{cm}^{-2} \text{Hz}^{-1}$). Since

$$\int_{\nu_1}^{\nu_2} \nu f_\nu d\nu = \int_{\ln \nu_1}^{\ln \nu_2} f_\nu d \ln \nu$$

⇒ plotting νf_ν in a log-log plot gives measure of energy emitted per frequency decade

Proportional Counters

Introduction

Before we can look at individual radiation processes, we need to understand how the radiation is detected:

- Non-imaging detectors

Detectors capable of detecting photons from a source, but without any spatial resolution

⇒ Require, e.g., **collimators to limit field of view.**

Example: Proportional Counters

- Imaging detectors


Detectors with a spatial resolution, typically used in the IR, optical, UV or for soft X-rays.

Generally behind some type of focusing optics.

Example: Charge coupled devices (CCDs)

Here: will concentrate on detectors for X-rays and optical light, starting with non-imaging detectors.

Ionization chamber



Simplest gas detector: gas-filled capacitor (**ionization chamber**)


- photon is photo absorbed in detector gas
- K-shell electron ejected with energy \propto initial E
- through collisional ionization N charges are produced
- Electrostatic induction induces charge on capacitor plates
- Charges drain via $R \Rightarrow$ Measurement

Pulse height: $\Delta Q = -Ne = C\Delta U \Rightarrow \Delta U = -Ne/C$

Typical magnitude of signal: $C = 20 \text{ pF}$, $Ne = 2 \times 10^5 e^- \cdot e \Rightarrow \Delta U = 1.6 \text{ mV}$.

Problem: Pulse very weak since only primary charge measured.

Proportional Counters



Solution: amplify charge

Close to Anode wire:

$E(r) = V/(r \ln(b/a))$ (b radius of cathode, a radius of anode)

⇒ Strong **acceleration** of ionized particles

⇒ **collisional ionization** of gas

⇒ **cascade!**

Measured voltage:


$$\Delta U = -\frac{eN}{C} \cdot A$$

where A : **amplification factor** (typical: $A = 10^4 \dots 10^6$).

Since $A \sim \text{const.}$: Voltage pulse $\propto N$, and therefore **Voltage pulse \propto detected X-ray energy!**
and therefore: **"proportional counter"**

Proportional counters

Pulse Amplification



(after Grupen, Fig.4.21)

Pulse amplification in detector as a function of anode voltage.

Typical proportional counter voltages are several 100 to 1000 V (depending on detector gas).

Proportional counters

Detector Gas

Use **inert gases**, e.g., Ar or Xe, since required voltage smallest and only low losses due to excitation of the gas atoms.

Number of ions produced: $N = E/\omega$, where ω is given by:

Gas	H	He	Ne	Ar	Kr
ω [eV]	36.6	44.4	36.8	26.25	24.1
Gas	Xe	Air	CO ₂	CH ₄	
ω [eV]	21.9	35.2	34.2	29.1	


⇒ Typically $N \sim 1000$ **electron-ion pairs** per 20 keV photon.

Note:

- probability for absorption $\sigma_{\text{bf}} \propto Z^{4\dots 5} \Rightarrow$ use **Xenon** ($Z = 54$) for astronomical detectors
- since $\sigma_{\text{bf}} \propto E^{-3} \Rightarrow$ **proportional counters limited to $E < 100$ keV.**

Proportional counters

Cascade, I




Gruppen (Fig. 4.9)

Typical electron paths $\approx \mu\text{m} \Rightarrow$ **Cascade happens very close to anode wire**

Proportional counters

Cascade, II



Gruppen (Fig. 4.27)

Electrons are accelerated to **very high speeds** towards wire, Ions are accelerated away from wire

⇒ **Main signal from ions**, *not* electrons, since ions have larger potential difference

typical duration of signal $\sim 100 \mu\text{s}$, can reach higher time resolution by differentiating the signal

Proportional counters

Energy Resolution

Measured signal: pulse height ⇒ Energy of the X-ray

Resolution: ΔE : Width (=FWHM, Full Width at Half Maximum) of the distribution of measured energies.

Poisson statistics (N discrete Electron-Ion pairs!):

$$\Delta E \propto 2.35\sqrt{N} \propto 2.35\sqrt{E}$$

Typically one uses $\Delta E/E$

Slight correlations due to amplifying discharge ⇒ Width of distribution somewhat smaller than expected from Poisson statistics ⇒ **Fano Factor F** :

$$\frac{\Delta E}{E} = 2.35 \left(\frac{F}{N} \right)^{1/2}$$

where for gas detectors $F \sim 0.2-0.3$

More detailed theory yields

$$\frac{\Delta E}{E} = 2.35 \left(\frac{W(F+A)}{E} \right)^{1/2}$$

W : mean energy to produce a pair (26 eV for Ar+Methane), $F \sim 0.2$, $A \sim 0.6$

⇒ up to 14% at 5.9 keV doable.

Proportional counters


Quenching

Problem: Excited ions emit UV photons \implies formation of new cascades due to photo effect \implies Total cascade takes a long time \implies **large dead time**.

Solution: Absorption of UV photons in “quenching gas”, which is added to primary photomultiplier gas \implies **cascade $< 1 \mu\text{s}$**

Energy of excited quenching gas is dumped later via inelastic collisions within the gas

Also direct quenching of cascade, e.g., via




Typical quenching gases: CH_4 , alcohol ($\text{C}_2\text{H}_5\text{OH}$), CO_2 , BF_3 , ... (about $\sim 10\%$ of total gas pressure).


Often used: “P10-gas” (90% Ar and 10% CH_4)

Proportional counters

Ageing



Marko Spegel, 1999 (Diss. Uni Wien), J. Vavra, SLAC-3882



Cascade: **plasma discharge** \implies Destruction of gas contaminants \implies formation of **free radicals** \implies **polymerization**


Polymers have high dipole moment \implies attach to electrodes \implies Reduction of pulse charges \implies “Ageing”
Typical contaminants: carbon, oxide-layers, silicates, e.g., from oil, finger grease, Silan (SiH_4), solvents in vacuum sealants, ...

Results in **field electron emission** through photo-effect \implies discharge \implies **wire destroyed** (“Malter-Effekt”)

Most sensitive proportional counter gas: Ar/ CH_4 ...

Proportional counters

Example: RXTE-PCA




Rossi-X-ray Timing Explorer, Launch 30.12.1995, 3 instruments:

- **Proportional Counter Array** (PCA, 2–100 keV),
- **High Energy X-ray Timing Experiment** (HEXTE, 15–250 keV),
- **All Sky Monitor** (ASM, 2–10 keV)

PCA and HEXTE have μsec timing resolution


Proportional counters

Example: RXTE-PCA



Proportional counters

Example: RXTE-PCA



PCA consists of 5 Proportional Counter Units (PCUs), 1 atm pressure

Propane and outer anodes in Anti-Coincidence \Rightarrow Background reduction


BeCu-Collimator to limit field of view to $\sim 1^\circ$

^{241}Am -source for energy calibration (simultaneous emission of 59.6 keV γ 's and α -particles \Rightarrow coincidence measurements).

Proportional counters

12

Example: RXTE-PCA



Response matrix: Relation between incident photon energy and detection channel.


Resolution: 18% at 6 keV

Escape-Peaks caused by Xe $K\alpha$ ($E = 29.46$ keV) and Xe $L\alpha$ ($E = 4.11$ keV) photons leaving the detector without being detected.

Proportional counters

14

Example: RXTE-PCA



... this is what happens once a Mylar window starts having a hole in space
 In addition **ageing** \implies Reduction of high voltage, alternate use of different PCUs.